



**Cooling Salve**

Covalent CC, LLC  
MME ID: NV20191018652  
Topical, Salve

Sample: DIGP2201.0034.T.00206

Sample Date: 01/14/2022 Report Date: 01/21/2022

METRC Sample:

Lot #: 1; Production Run #: SLV2112S2000C;

**AMENDED**

**Potency Test Results**

**Cannabinoid Test Results**

**Terpene Test Results**

**Not Tested**

	<b>&lt;LOQ</b>	<b>2,741.700</b>
	Total Potential THC	Total Potential CBD
	<b>&lt;LOQ</b>	<b>&lt;LOQ</b>
	THCa	CBDa

Analyte	CAS No.	LOQ	Mass	Mass
---------	---------	-----	------	------

1 Unit = 1 unit, 57g

Analyte	LOQ	Mass	Mass
	%	mg/unit	%
THCa	0.0100	<LOQ	<LOQ
Δ9-THC	0.0100	<LOQ	<LOQ
Δ8-THC	0.0100	<LOQ	<LOQ
THCV	0.0100	<LOQ	<LOQ
CBDa	0.0100	<LOQ	<LOQ
CBD	0.0100	2741.700	4.8100
CBDV	0.0100	16.815	0.0295
CBN	0.0100	<LOQ	<LOQ
CBGa	0.0100	<LOQ	<LOQ
CBG	0.0100	<LOQ	<LOQ
CBC	0.0100	<LOQ	<LOQ
<b>Total</b>		<b>2758.515</b>	<b>4.8395</b>

Total Potential THC = (THCa \* 0.877) + Δ9-THC + Δ8-THC, Total Potential CBD = (CBDa \* 0.877) + CBD, LOQ = Limit of Quantitation; NR = Not Reported; ND = Not Detected; Cannabinoids for flower and trim reported as received. Cannabinoids analyzed per Digipath Labs SOP-317 on an Agilent 1260 UPLC.

NR = Not Reported; ND = Not Detected; LOQ = Limit of Quantification. Terpenes analyzed per Digipath Labs SOP-333 on an Agilent FID-GCMS or Digipath Labs SOP-334 on an Agilent 7697A/7890B/5977A Headspace GC/MS.

**Safety & Quality Tests**

<b>Visual</b>	Not Tested	<b>Moisture Content</b>	Not Tested
<b>Microbiological</b>	Not Tested	<b>Gender</b>	Not Tested
<b>Heavy Metals</b>	Not Tested	<b>Residual Solvents</b>	Not Tested
<b>Mycotoxins</b>	Not Tested	<b>Pesticides</b>	Not Tested
<b>Water Activity</b>	Not Tested		

**CANNABINOIDS LABELING SUGGESTION PER NAC 453A.508**

Total Potential THC	CBN	Total Potential CBD
<LOQ	<LOQ	2,741.700 mg/unit

Scan to View Results



I certify that this sample has been tested by DigiPath Labs. All results are reported on AS-IS basis.

Unit Weight correct from 2000g to 57g.

*Sherrif Defreece*

Sherrif Defreece  
Scientific Director



Accreditation #99721

All pass/fail limits are as specified in nCCR, NAC 453.A and DPBH Policies. Unless otherwise stated, all quality control samples performed within specifications previously established by the Laboratory. This product has been tested by Digipath Labs, Inc. using validated testing methodologies under a QMS as required by ISO-17025:2017 and Nevada state law. Sample collected per Digipath Labs' SOP-312. Values reported relate only to the product tested. Water activity tested at 25o C. Digipath Labs, Inc. makes no claims as to the efficacy, safety or other risks associated with any detected or non-detected levels of any compounds reported herein and treats all client communication and testing results as confidential. This Certificate shall not be reproduced without the written approval of Digipath Labs, Inc. Measurement Uncertainty values have been determined for all methods and analytes.