

## Exo-THC Analysis Advice

**Manufacturer:** Covalent CC, LLC (dba Covalent Custom Cannabinoids)

**FDA Food Facility Registration Number:** 17178984546

**Registration Expiration Date:** 12/31/2022

**Date:** January 1, 2022



To whom it may concern,

We have recently discovered unexpected false positive lab results for our THC Free Broad Spectrum Crystal Resistant Distillate. When submitting raw material samples to labs for analysis, depending on if it's a standard they offer - labs will often report false positives for THC due to the false peaks the **Exo-THC** can create. The best practice is to inform lab that the material is THC-free material that is also *extremely unique in composition*, with surprisingly high levels of minor cannabinoids that are very uncommon for labs to see, and therefore create various challenges to determine accurate results in comparison to the status-quo cannabinoid potency analysis. What has been helpful in the past to produce accurate results of "ND" THCa/Delta 9/Delta 8 is to **specifically request for Exo-THC to be tested on the material**.



**Fred Palmer**  
Regulatory Compliance Manager  
[fp@covalentcbd.com](mailto:fp@covalentcbd.com)

